

Neuron MAX 088

Soft Tip, 6 F Long Sheath
for Neurovascular Support

ACE
Reperfusion
Catheter

Long Sheath Designed for Neurovascular
Use Delivering MAX Support for
Today's Advanced Therapies

Neuron
MAX 088


Penumbra 

Neuron MAX 088 Soft Tip, 6 F Long Sheath

- Full length stainless steel braid reinforced proximal support zone transitioning to 4 cm distal flexible zone with large .088" (2.24 mm) inner lumen
- Distal shaft has hydrophilic coating designed for optimal trackability
- Available in 80 cm, 90 cm, and 100 cm lengths with Straight and Multipurpose (MP) tip shapes

6 F Select Catheter

1. Enables primary access and atraumatic placement into desired vessel without an over-the-wire exchange
2. Useful for diagnostic angiogram – .040" (1.02 mm) lumen
3. Stainless steel braided proximal shaft for support and torque response with a soft polymer distal shaft to enable atraumatic vessel selection



Ordering Information

Neuron MAX 6 F 088 Lumen Long Sheath (Crosscut Valve, RHV and Dilator included)

Catalog Number	Description	Tip Shape	Working Length	Distal Flexible Zone	Outer Diameter* Proximal / Distal	Inner Diameter	Wire Compatibility
PNML6F088804	6 F 088 Neuron MAX Long Sheath, 80/4 Straight		80 cm	4 cm	8 F / 8 F	.088" (2.24 mm)	.035/.038" (.89 mm/.97 mm)
PNML6F088804M	6 F 088 Neuron MAX Long Sheath, 80/4 MP		80 cm	4 cm	8 F / 8 F	.088" (2.24 mm)	.035/.038" (.89 mm/.97 mm)
PNML6F088904	6 F 088 Neuron MAX Long Sheath, 90/4 Straight		90 cm	4 cm	8 F / 8 F	.088" (2.24 mm)	.035/.038" (.89 mm/.97 mm)
PNML6F088904M	6 F 088 Neuron MAX Long Sheath, 90/4 MP		90 cm	4 cm	8 F / 8 F	.088" (2.24 mm)	.035/.038" (.89 mm/.97 mm)
PNML6F0881004	6 F 088 Neuron MAX Long Sheath, 100/4 Straight		100 cm	4 cm	8 F / 8 F	.088" (2.24 mm)	.035/.038" (.89 mm/.97 mm)
PNML6F0881004M	6 F 088 Neuron MAX Long Sheath, 100/4 MP		100 cm	4 cm	8 F / 8 F	.088" (2.24 mm)	.035/.038" (.89 mm/.97 mm)

6 F Select Catheter

Catalog Number	Description	Tip Shape	Working Length	Distal Flexible Zone	Outer Diameter* Proximal / Distal	Inner Diameter	Wire Compatibility
PNS6F105BER	6 F Select Catheter, 105 BER		105 cm	9 cm	5.6 F / 6 F / 5 F	.040" (1.02 mm)	.035/.038" (.89 mm/.97 mm)
PNS6F105H1	6 F Select Catheter, 105 H1		105 cm	9 cm	5.6 F / 6 F / 5 F	.040" (1.02 mm)	.035/.038" (.89 mm/.97 mm)
PNS6F125BER	6 F Select Catheter, 125 BER		125 cm	9 cm	5.6 F / 6 F / 5 F	.040" (1.02 mm)	.035/.038" (.89 mm/.97 mm)
PNS6F125SIM	6 F Select Catheter, 125 SIM		125 cm	9 cm	5.6 F / 6 F / 5 F	.040" (1.02 mm)	.035/.038" (.89 mm/.97 mm)
PNS6F125H1	6 F Select Catheter, 125 H1		125 cm	9 cm	5.6 F / 6 F / 5 F	.040" (1.02 mm)	.035/.038" (.89 mm/.97 mm)
PNS6F125SIMV	6 F Select Catheter, 125 SIM-V		125 cm	9 cm	5.6 F / 6 F / 5 F	.040" (1.02 mm)	.035/.038" (.89 mm/.97 mm)

* A mid outer diameter is only listed if applicable to device.

Prior to use, please refer to the Instructions for Use for complete product indications, contraindications, warnings, precautions, potential adverse events, and detailed instructions for use.

Penumbra System – Intended Use

The Penumbra System is intended for use in the revascularization of patients with acute ischemic stroke secondary to intracranial large vessel occlusive disease using continuous aspiration.

Potential Adverse Events

Possible complications include, but are not limited to, the following: allergic reaction and anaphylaxis from contrast media; acute occlusion; air embolism; arteriovenous fistula; death; device malfunction; distal embolization; emboli; false aneurysm formation; hematoma or hemorrhage at access site; inability to completely remove thrombus; infection; intracranial hemorrhage; ischemia; kidney damage from contrast media; neurological deficits including stroke; vessel spasm, thrombosis, dissection, or perforation.

Neuron MAX System – Intended Use

The Neuron MAX System is intended for the introduction of interventional devices into the peripheral, coronary, and neuro vasculature.

Potential Adverse Events

Possible complications include, but are not limited to, the following: acute occlusion; air embolism; death; distal embolization; emboli; false aneurysm formation; hematoma or hemorrhage at puncture site; infection; intracranial hemorrhage; ischemia; neurological deficits including stroke; vessel spasm, thrombosis, dissection, or perforation.

Penumbra



Penumbra, Inc. USA
One Penumbra Place
Alameda, CA 94502
USA
1.888.272.4606
T 1.510.748.3200
F 1.510.748.3232
order@penumbrainc.com
info@penumbrainc.com

Penumbra Europe GmbH
Am Borsigturm 44
13507 Berlin
Germany
T +49 30 2005 676-0
F +49 30 2005 676-10
de-order@penumbrainc.com
info@penumbrainc.de

Penumbra Neuro Australia Pty Ltd
Suite 3, Level 5, 1 Oxford Street
Darlinghurst NSW 2010
Australia
T +61-1300 817 025
F +61-1300 817 026
order.anz@penumbrainc.com

Penumbra Latin America
Distribuidora de Equipamentos e
Produtos Médicos Ltda
Avenida Brigadeiro Luis Antonio
3421 cj 201 CEP 1401-001
Sao Paulo, Brazil
T 5511.2883.5825
order.la@penumbrainc.com

Product availability varies by country. Please contact your local Penumbra representative for more information.

Copyright ©2019 Penumbra, Inc. All rights reserved. The Penumbra P logo, Neuron, Penumbra System, ACE, Select, and MAX are registered trademarks or trademarks of Penumbra, Inc. in the USA and other countries. All other trademarks are the property of their respective owners.

15581, Rev. B 06/19 OUS